


**Highlights:**

- Rapid self-propelled micromotors-based immunosensing strategy for cortisol.
- Naked-eye detection of cortisol in 2 min using the TMB/H<sub>2</sub>O<sub>2</sub> system.
- Greatly enhanced speed and sensitivity using antibody-modified micromotors.
- “On the move” efficient binding of 0.1 µg mL<sup>-1</sup> cortisol using microliter samples.
- Concept readily translated to the detection of a broad range of target biomarkers.

## Graphical Abstract



Naked-eye cortisol sensing based on receptor-functionalized micromotors.

# Rapid Micromotor-Based Naked-Eye Immunoassay

*Berta Esteban-Fernández de Ávila,<sup>†</sup> Mingjiao Zhao,<sup>†, ‡</sup> Susana Campuzano,<sup>§</sup> Francesco Ricci,<sup>‡</sup> José M. Pingarrón,<sup>§</sup> Marcello Mascini,<sup>\*</sup> <sup>†, +, §</sup> and Joseph Wang<sup>\*</sup> <sup>†</sup>*

<sup>†</sup> Department of Nanoengineering, University of California, San Diego, La Jolla, California 92093, United States

<sup>‡</sup> Department of Chemistry, University of Rome, Tor Vergata, Via della Ricerca Scientifica, 00133, Rome, Italy

<sup>§</sup> Department of Analytical Chemistry, Complutense University of Madrid, E-28040 Madrid, Spain

<sup>+</sup> Faculty of Bioscience and Technology for Food, Agriculture and Environment, University of Teramo, 64100, Teramo, Italy

## Corresponding Author

\*E-mail address: [josephwang@ucsd.edu](mailto:josephwang@ucsd.edu), [mmascini@unite.it](mailto:mmascini@unite.it)

**ABSTRACT:** A dynamic micromotor-based immunoassay, exemplified by cortisol detection, based on the use of tubular micromotors functionalized with a specific antibody is described. The use of antibody-functionalized micromotors offers huge acceleration of both direct and competitive cortisol immunoassays, along with greatly enhanced sensitivity of direct and competitive immunoassays. The dramatically improved speed and sensitivity reflect the greatly increased likelihood of antibody-cortisol contacts and fluid mixing associated with the dynamic movement of these microtube motors and corresponding bubble generation that lead to a highly efficient and rapid recognition process. Rapid naked-eye detection of cortisol in the sample is achieved in connection to use of horseradish peroxidase (HRP) tag and TMB/H<sub>2</sub>O<sub>2</sub> system. Key parameters of the competitive immunoassay (e.g., incubation time and reaction volume) were optimized. This fast visual micromotor-based sensing approach enables “on the move” specific detection of the target cortisol down to 0.1 µg mL<sup>-1</sup> in just 2 min, using ultrasmall (50 µL) sample volumes.

## Keywords:

Cortisol, real-time sensing, micromotor, antibody, naked-eye detection

## 1. Introduction

Recent advances in nanomotors have paved the way to novel biosensing systems and applications. Particularly, the ability of nano/microscale motors to capture and transport specific target analytes in complex biological matrixes has been extensively explored [1-4].

Nano/micromotors functionalized with specific bioreceptors have thus allowed “on-the-fly” detection of different bioanalytes, ranging from nucleic acids to tumor cells [5-9]. Tubular micromotors functionalized with specific antibodies have been incorporated into lab-on-a-chip diagnostic devices demonstrating efficient recognition and isolation of the model protein IgG [10, 11]. Such direct “on-the-fly” recognition and isolation processes obviate the need for common sampling and washing procedures. The efficient movement of these micromotors and generated-microbubbles tail has been shown to enhance the fluid mixing of the sample and to greatly increase the analyte–receptor interactions toward faster and more sensitive assays [12-14]. These dynamic sensing nanosystems have demonstrated important advantages to enhance target-bioreceptor interactions, which hold considerable interest for its implementation as medical diagnostic tools.

In this communication, we report on a new nanomotor immunoassay strategy based on naked-eye detection system. Cortisol, an important stress and clinical biomarker [15-19], was selected as model target to demonstrate the capabilities of this new approach. The most widely used methods in clinical assays of cortisol include luminescence and fluorescence assays [20, 21], enzyme-linked immunosorbent assays (ELISAs) [22], and immunoassays in connection to fluorescence [23] or electrochemical detection [17, 24-28]. However, these cortisol-detection methods still present some limitations in terms of analysis time, pre-treatment steps, and cost and do not involve rapid naked-eye detection.

The new dynamic cortisol detection strategy, described in this paper, is based on the use of tubular micromotors functionalized with a specific anti-cortisol antibody (Figure 1). The advantages of such micromotor-based cortisol naked-eye detection system have been illustrated using direct and competitive immunoassay formats. As illustrated in Figure 1, the direct motor-based immunoassay relies on HRP-labeled cortisol. The anti-cortisol functionalized-micromotors efficiently swim through the sample and detect rapidly and specifically the tagged-cortisol target (Figure 1A). The competitive motor assay involves

1 competition of free-cortisol and the cortisol-HRP for the antibody sites on the moving  
2 micromotors. In both assays, short incubation of the resulting cortisol-HRP modified  
3 micromotors in a TMB/H<sub>2</sub>O<sub>2</sub> solution offers absorbance measurements of the deep blue  
4 product of TMB at 690 nm (Figure 1B). The absorbance spectra of samples prepared in the  
5 absence and presence of cortisol-HRP (light and dark blue lines, corresponding to transparent  
6 and blue color solutions, respectively) are displayed in the top part of Figure 1B.

7  
8  
9  
10  
11 The combination of the specific antibody with the efficient movement of the H<sub>2</sub>O<sub>2</sub>-  
12 propelled tubular micromotors allows rapid detection of dilute amounts of cortisol-HRP using  
13 microliter sample volumes. In comparison with early micromotor-based immunosensing  
14 strategies, involving direct or sandwich approaches and immunoreagents tagged with  
15 microsphere tracers for visualizing the antibody-antigen interaction [10, 11], this micromotor  
16 strategy demonstrated for the first time the use of a direct and competitive immunoassays with  
17 rapid naked-eye detection of the target analyte. As will be illustrated in the following sections,  
18 the movement of the antibody-modified micromotors through the sample plays a critical role  
19 in the enhanced biosensing performance. Such continuous motion of multiple anti-cortisol-  
20 micromotors imparts effective sample mixing within microliter samples that addresses the  
21 limitations associated with the slow cortisol transport under quiescent conditions and leads to  
22 fast antibody-cortisol interactions and hence to higher speed and sensitivity.

## 33 **FIGURE 1**

## 36 **2. Materials and methods**

### 38 **2.1. Reagents**

39  
40 A Cyclopore polycarbonate (PC) membrane, containing 5  $\mu\text{m}$  diameter conical-shaped  
41 micropores (Catalog No 110613; Whatman, Maidstone, U.K.), was employed as the template.  
42 Hydrocortisone (H4001, MW 362.46 g mol<sup>-1</sup>), Sodium dodecyl sulfate (SDS, MW 288.38 g  
43 mol<sup>-1</sup>), 3,4-ethylenedioxythiophene (EDOT), potassium nitrate (KNO<sub>3</sub>), hydrogen peroxide  
44 (H<sub>2</sub>O<sub>2</sub>), 11-Mercaptoundecanoic acid (MUA), 6-Mercapto-1-hexanol (MCH), 4-  
45 morpholineethanesulfonic acid hemisodium salt (MES), 1-ethyl-3-[3-dimethylaminopropyl]  
46 carbodiimide hydrochloride (EDC), and N-hydroxysulfosuccinimide (Sulfo-NHS) were  
47 purchased from Sigma-Aldrich<sup>®</sup>. The monoclonal cortisol antibody (Catalog # P01-92-94M-  
48 P) and cortisol-HRP antigen (Catalog # P91-92-91H) were purchased from EastCoast Bio<sup>®</sup>.  
49 K-Blue<sup>®</sup> Aqueous 3,3',5,5'-tetramethylbenzidine (TMB) substrate was purchased from  
50 Neogen<sup>®</sup> Corporation.

1 The antibody was dissolved in 0.05 M phosphate buffer (PBS) pH 7.4. The washing  
2 solution consisted of 0.05 M phosphate buffer pH 7.4 containing 0.05% Tween-20.

3 The aqueous hydrogen peroxide solutions were used as the chemical fuel, and sodium  
4 dodecyl sulfate (SDS, MW 288.38 g mol<sup>-1</sup>) (at 3% final concentration) was used as a  
5 surfactant in all propulsion experiments.  
6

7  
8  
9 Ultrapure water (18MX cm, Millipore Corporation, USA) was used for the preparation  
10 of all aqueous solutions  
11

## 12 13 14 **2.2. Apparatus** 15

16 Template electrochemical deposition of multi-layer microtubes was carried out with a  
17 CHI 661D potentiostat (CH Instruments, Austin, TX). Scanning electron microscopy (SEM)  
18 images were obtained with a Phillips XL30 ESEM instrument, using an acceleration voltage  
19 of 20 kV. Energy-dispersive X-ray mapping analysis was performed using an Oxford EDX  
20 detector attached to SEM instrument and operated by INCA software. An inverted optical  
21 microscope (Nikon Instrument Inc. Ti-S/L100), coupled with a 20× and 40× objectives, a  
22 Photometrics QuantEM 512/SC camera (Roper Scientific, Duluth, GA) and MetaMorph 7.6  
23 software (Molecular Devices, Sunnyvale, CA) were used for capturing movies at a frame rate  
24 of 30 frames per sec. The speed of the micromotors was tracked using a Metamorph tracking  
25 module and the results were statistically analyzed using Origin software.  
26

27 The absorbance was measured between 500 and 800 nm using a UV-2450 Shimadzu  
28 spectrophotometer.  
29  
30  
31  
32  
33

## 34 35 36 37 38 39 **2.3. Preparation of tubular micromotors** 40

41 The tubular micromotors were prepared using a common template-directed  
42 electrodeposition protocol. A Cyclopore polycarbonate membrane, containing 5 μm diameter  
43 conical-shaped micropores (Catalog No 110613; Whatman, Maidstone, U.K.), was employed  
44 as the template. A 75 nm-thick gold film was first sputtered on one side of the porous  
45 membrane to serve as the working electrode using the Denton Discovery 18 sputtering  
46 system. The coating was performed at room temperature under base vacuum of 5×10<sup>-6</sup> Torr,  
47 DC power 200 W and Ar pressure of 3.1 mT, along with a rotation speed of 65 rpm and  
48 sputtering time of 90 s. A Pt wire and Ag/AgCl (3 M KCl) served as counter and reference  
49 electrodes, respectively. The membrane was then assembled in the electrochemical plating  
50 cell with an aluminum foil serving as a contact. Poly(3,4-ethylenedioxythiophene) (PEDOT)  
51 microtubes were electropolymerized at +0.80 V using a charge of 0.1 C from a plating  
52  
53  
54  
55  
56  
57  
58  
59  
60  
61  
62  
63  
64  
65

1 solution containing 15 mM EDOT, 7.5 mM KNO<sub>3</sub> and 100 mM SDS; subsequently, Ni was  
2 deposited from a Ni plating solution containing NiCl<sub>2</sub> · 3.6H<sub>2</sub>O (20 g L<sup>-1</sup>), Ni(H<sub>2</sub>NSO<sub>3</sub>)<sub>2</sub> · 4H<sub>2</sub>O  
3 (515 g L<sup>-1</sup>), and H<sub>3</sub>BO<sub>3</sub> (20 g L<sup>-1</sup>) at -1.3 V (*vs* Ag/AgCl) for 4 °C. The inner Pt layer was  
4 galvanostatically deposited at -2 mA for 500 s from a commercial platinum plating solution  
5 (Platinum RTP; Technic Inc, Anaheim, CA). The sputtered gold layer was completely  
6 removed by hand polishing with alumina slurry (3–4 μm). After that, the membrane was  
7 dissolved in methylene chloride for 30 min to completely release the microtubes. The  
8 micromotors were collected by centrifugation at 7,000 rpm for 3 min and washed with  
9 methylene chloride, followed by ethanol and ultrapure water (18.2 MΩ cm), three times for  
10 each, with a 3 min centrifugation following each wash. The template preparation method  
11 resulted in decidedly reproducible micromotors. The micromotors were placed over a glass  
12 slide and were sputtered with a layer of Au (DC power 200 W, Ar pressure of 2.9 mT, along  
13 with a rotation speed of 65 rpm and sputtering time of 40 s) for the following  
14 functionalization with thiols chemistry. Finally, the Au-sputtered micromotors were released  
15 from the glass slide by softly mechanical scratching.  
16  
17

18 Au-sputtered PEDOT/Ni microtubes were used for static control experiments,  
19 following the same fabrication protocol but without the electrodeposition of platinum, which  
20 is responsible of the decomposition of H<sub>2</sub>O<sub>2</sub> and consequent bubble production from the  
21 tubular micromotor.  
22  
23  
24  
25  
26  
27

## 28 **2.4. Micromotor functionalization**

29 The outer Au layer of the tubular micromotors or the static microtubes was  
30 functionalized first with a mixed MUA/MCH alkanethiol monolayer. A solution mixture of  
31 2.5 mM MUA and 7.5 mM MCH was prepared in ethanol. The micromotors or microtubes  
32 were incubated in the MUA/MCH solution overnight. After washing the micromotors or  
33 microtubes with water for 5 min, they were transferred to an eppendorf vial containing a 100  
34 μL MES buffer (pH 6.5) solution containing the coupling agents EDC and Sulfo-NHS at 0.4  
35 and 0.1 M, respectively, and were incubated for 35 min. After washing with PBS buffer pH  
36 7.4 for 5 min, the micromotors or microtubes were incubated with the anti-cortisol antibody  
37 (100 μg mL<sup>-1</sup>) during 2 h at room temperature, followed by two 15 min washing steps with a  
38 PBS solution (pH 7.4) containing 0.05 wt % tween-20 and one washing step with a PBS  
39 solution (pH 7.4) to remove the excess of antibody from the solution. The micromotors or  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60  
61  
62  
63  
64  
65

1 microtubes were centrifuged and kept dried at 4 °C until use. A schematic illustration of the  
2 protocol used for anti-cortisol covalent immobilization is shown in Fig.S1.

## 3 **2.5. Detection of attached cortisol-HRP**

4  
5 The antibody-functionalized micromotors were mixed with a PBS (pH 7.4) solution  
6 containing 1% H<sub>2</sub>O<sub>2</sub> fuel, 3% SDS surfactant and 1:100 of the cortisol-HRP target (all these  
7 variables were kept constant in all the experiments). The mixture was allowed to react during  
8 different times (30 s, 1, 2, 5 or 10 min) and also prepared in different volumes (25, 50, and  
9 100 µL). Afterwards, the micromotors were washed two times with PBS solution (pH 7.4)  
10 containing 0.05 wt % tween-20 and another time with PBS (pH 7.4) to remove the excess of  
11 unbounded cortisol-HRP. The cortisol-HRP-antibody-functionalized micromotors were  
12 separated by centrifugation and were mixed with 500 µL of undiluted TMB solution  
13 (containing H<sub>2</sub>O<sub>2</sub>), and the absorbance spectra (500–800nm) was measured. Notice that the  
14 TMB develops a deep blue color in the presence of HRP and H<sub>2</sub>O<sub>2</sub>. A schematic illustration  
15 of the protocol used for cortisol-HRP detection is shown in Fig.S1.

16  
17 Since the micromotors used were propelled by the bubbles generated at the inner Pt  
18 surface and the H<sub>2</sub>O<sub>2</sub> is required for TMB developing the blue color in the presence of HRP,  
19 anti-cortisol-functionalized-Au-sputtered PEDOT/Ni microtubes (without Pt inner layer) were  
20 used to perform the static micromotors control experiments.

## 21 **2.6. Competitive immunoassay**

22  
23 The direct competitive immunoassay for cortisol detection was performed by mixing  
24 antibody-functionalized micromotors or microtubes (for static control experiments) with 50  
25 µL of a PBS (pH 7.4) solution containing 1% H<sub>2</sub>O<sub>2</sub> fuel, 3% SDS surfactant, different  
26 concentrations of a free cortisol standard (0, 0.1 and 1 µg mL<sup>-1</sup>) and 0.5 µL of the commercial  
27 cortisol-HRP target solution (i.e., 1:100 dilution). After 2 and 10 min incubation when using  
28 micromotors and stationary microtubes, respectively, they were washed one time with PBS  
29 solution (pH 7.4) containing 0.05 wt % tween-20 and another time with PBS (pH 7.4).  
30 Finally, the micromotors and microtubes were separated by centrifugation and were mixed  
31 with 500 µL TMB/H<sub>2</sub>O<sub>2</sub> undiluted commercial solution, and the corresponding absorbance  
32 spectra and value at 690 nm were measured.



### 3. Results and discussion

#### 3.1. Synthesis and characterization of micromotors

Fig. 2A schematically illustrates the steps involved in the fabrication and functionalization of the multilayer PEDOT/Ni/Pt tubular micromotors with the anti-cortisol antibody. The micromotors used for detecting cortisol were fabricated by a standard membrane template electrodeposition method. Briefly, the template-electrodeposition protocol consisted on electropolymerization of the EDOT monomer within the conical micropores of a polycarbonate (PC) membrane used as a template, followed by electrodeposition of Ni and the inner Pt layer (Fig. 2A (a)), and release of the resulting PEDOT/Ni/Pt micromotors by dissolving the membrane (Fig. 2A (b)). The micromotors were subsequently sputtered with a thin layer of Au (Fig. 2A (c)) for further functionalization. The immobilization of the anti-cortisol antibody onto the outer Au layer of the micromotors was carried out using EDC/NHS chemistry through a mixed self-assembled monolayer (SAM) of MUA/MCH alkanethiols (Fig. 2A (d); detailed experimental protocol explained in Methods and in Fig.S1). As was mentioned above, Au-sputtered PEDOT/Ni microtubes were used as static controls, following the same fabrication protocol but without the Pt electrodeposition, which is responsible for the catalytic  $\text{H}_2\text{O}_2$  fuel reaction and consequent bubble production from the tubular micromotor.

#### **FIGURE 2**

Fig. 2B depicts the scanning electron microscopy (SEM) image of an Au-sputtered PEDOT/Ni/Pt micromotor with  $\sim 10\ \mu\text{m}$  length and a  $\sim 5\ \mu\text{m}$  diameter; the latter reflects the pore size of the PC membrane used in the fabrication process. Corresponding energy-dispersive X-ray (EDX) images of the Au-sputtered PEDOT/Ni/Pt micromotor confirmed the presence of the outer Au film, Ni intermediate layer, and an inner Pt layer (Fig. 2B), required for antibody functionalization, concentration and magnetic guidance, and catalytic propulsion in the presence of  $\text{H}_2\text{O}_2$ , respectively.

The propulsion of the antibody-functionalized Au-sputtered PEDOT/Ni/Pt micromotors was also studied. Time-lapse images, taken from video S1, illustrate the propulsion of the antibody-functionalized micromotors in the presence of 1%  $\text{H}_2\text{O}_2$  fuel, along with the magnetic guidance of the micromotor through the cortisol spiked-buffer (Fig. 2C). Efficient movement at a speed of  $80\ \mu\text{m s}^{-1}$  is observed. The inclusion of the Ni layer in the

composition of the micromotor facilitates the magnetic manipulation of the micromotors during the functionalization protocol, and opens the possibility of guiding these antibody-functionalized micromotors to the target destination.

### 3.2. Optimization of detection conditions

The development of the immunosensing nanomotor strategy has involved optimization of key experimental conditions, such as incubation time and reaction volume, and variety of relevant control experiments (Fig. 3).

Initial control experiments were performed by incubating anti-cortisol-micromotors in a blank solution (prepared in the absence of cortisol-HRP), or non-functionalized micromotors in a cortisol-HRP solution (Fig. 3A: bars 1 and 2, respectively). These resulted in negligible absorbance, thus indicating the absence of unspecific background contributions due to non-specific adsorptions of the enzymatic tracer onto the non-modified micromotors and discarding potential erroneous readings.

In order to demonstrate the important role of the self-propelled micromotors, and corresponding localized fluid convection, in the acceleration and efficiency of the cortisol binding and detection, the effect of the incubation time was investigated both under static and moving conditions, using anti-cortisol functionalized microtubes and micromotors, respectively. As shown in Fig. 3, a clear increase of the signal is observed when using dynamic anti-cortisol-micromotors compared to static anti-cortisol-microtubes. In particular, moving the micromotors for 30 s provided a 3-fold higher absorbance signal compared to static microtubes after 1 min incubation. This trend was confirmed by the double response provided by the dynamic micromotors at 1 min compared to that of the static microtubes after 5 min. Due to saturation of the absorbance above 5 min, this was the longest time tested. Overall, the data of Figure 3A confirm that the rapid movement of the anti-cortisol-micromotors through the sample greatly increases the likelihood of antibody-cortisol contacts, and thus leads to a highly efficient and rapid recognition process.

### FIGURE 3

After studying the cortisol-detection ability of the anti-cortisol-micromotors at different incubation times, the effect of the reaction volume (25, 50 and 100  $\mu$ L) was investigated using a fixed reaction time of 2 min. Figure 3B demonstrates that the assay sensitivity increases substantially upon decreasing the sample volumes from 100 to 25  $\mu$ L, reflecting the greatly enhanced likelihood of antibody-cortisol contacts using smaller sample volumes under these dynamic conditions. Overall, these preliminary experiments confirm that

the use of dynamic micromotors clearly improves the antibody-cortisol interactions, to offer increasing sensitivity while reducing largely the assay time.

### 3.3. Direct competitive assay

In addition to direct immunoassays of cortisol, the micromotor approach can greatly enhance the performance of competitive cortisol immunoassays (Fig. 4).

#### FIGURE 4

These bioassays involve the competition of cortisol and HRP-tagged cortisol for the antibody sites on the moving micromotors (Fig. 4A). The direct competitive immunoassay format was carried out using both static and moving antibody-modified microtubes and micromotors to evaluate their recognition properties in a semi-quantitative manner. Using this approach, the quantity of cortisol-HRP, inversely proportional to the quantity of free cortisol in the sample, was rapidly estimated.

According to the results obtained from the previous optimization studies, the reaction volume selected for performing this competitive immunoassay was 50  $\mu\text{L}$ , using reaction times of 2 and 10 min for dynamic and static conditions, respectively. The anti-cortisol functionalized static microtubes and micromotors were incubated in standard solutions prepared in the absence and in the presence of 0.1 and 1  $\mu\text{g mL}^{-1}$  of free cortisol. As demonstrated results presented in Figs. 4B and 4C, absorbance signals for cortisol-HRP detected under static and dynamic conditions decreased with increasing the free cortisol concentration, in agreement with the basis of a direct competitive immunoassay.

In order to evaluate the influence of the free cortisol binding efficiency upon the micromotors movement, competition plots obtained under moving and static conditions were compared. As it is shown in Figs. 4D and 4E, the moving micromotors offer a drastically rise in the negative slope of dynamic system response (Fig. 4E), nearly 5-fold more than their static counterpart in the range 0-0.1  $\mu\text{g mL}^{-1}$  free cortisol concentration (Fig. 4D). In agreement with previous results obtained in the optimization of the incubation time, these data confirm the possibility of using these anti-cortisol functionalized micromotors to accelerate the selective antibody-target interaction while increasing the assay sensitivity and performance. These preliminary results demonstrate that the micromotor approach offers both the qualitative naked-eye detection of lateral flow assays (the most practical approaches to point of care cortisol testing) and the possibility to perform also quantitative analysis.

Overall, the presented results demonstrate that the combination of specific anti-cortisol antibodies with the efficient movement of self-propelled micromotors allows rapid detection of dilute amounts of cortisol-HRP in microliter volume samples. The attractive capabilities demonstrated by these functionalized micromotors for rapid and effective cortisol sensing hold enormous potential for designing next-generation of miniaturized mobile diagnostic devices.

## 4. Conclusions

We have developed a very attractive micromotor-based immunosensing strategy for rapid, selective, single-step and naked-eye detection of cortisol. The new direct and competitive immunoassays exploit the transport properties of antibody-modified self-propelled micromotors for greatly enhanced speed and sensitivity in connection to an HRP tag for such naked-eye detection. Due to the increased localized fluid convection produced by their autonomous moving, the use of self-propelled antibody functionalized micromotors offers efficient target binding using microliter samples, and results in dramatic acceleration compared to antibody functionalized static microtubes. Such naked-eye detection greatly simplifies the whole immunoassay protocol. The improved capabilities of this “on-the-flight” direct immunoassay allowed sensitive ( $0.1 \mu\text{g L}^{-1}$ ) and rapid (2 min) detection of cortisol in ultrasmall sample volumes. It is worth to mention also that the attractive capabilities offered by these moving micromotors are advantageous compared to that of commercial micromagnetic counterparts in bioassays implementation in terms of simplicity of operation and reaction time. The magnetic guidance, associated with its intermediate Ni layer, could facilitate movement through a microchannel network toward a cortisol microchip.

These results demonstrate that the dynamic cortisol detection strategy, based on the use of tubular micromotors functionalized with a specific anti-cortisol antibody, can offer significant improvements and facilitate high-performance to direct and competitive immunoassays of cortisol. These rapid micromotors cortisol assays could be used also for rapid screening purposes in connection to existing cortisol detection systems that demand more time and laborious protocols. While the new concept was presented in connection to the detection of cortisol, it may be readily translated to the detection of a broad range of target biomarkers, making these micromotors extremely attractive biosensing tools.

## Acknowledgement

This work was supported by the Defense Threat Reduction Agency Joint Science and Technology Office for Chemical and Biological Defense (grants nos. HDTRA1-13-1-0002 and HDTRA1-14-1-0064), and by the European Union's Horizon 2020 Research and Innovation programme under the Marie Skłodowska-Curie grant agreement No 656687.


## Appendix A. Supplementary material

Supplementary data associated with this article can be found in the online version.


## References

- [1] J. Wang, *Nanomachines: fundamentals and applications*, John Wiley & Sons 2013.
- [2] S. Sánchez, L. Soler, J. Katuri, *Chemically Powered Micro- and Nanomotors*, *Angewandte Chemie International Edition* 54(5) (2015) 1414-1444.
- [3] M. Guix, C.C. Mayorga-Martinez, A. Merkoçi, *Nano/micromotors in (bio) chemical science applications*, *Chemical reviews* 114(12) (2014) 6285-6322.
- [4] H. Wang, M. Pumera, *Fabrication of micro/nanoscale motors*, *Chemical reviews* 115(16) (2015) 8704-8735.
- [5] X. Yu, Y. Li, J. Wu, H. Ju, *Motor-based autonomous microsensor for motion and counting immunoassay of cancer biomarker*, *Analytical chemistry* 86(9) (2014) 4501-4507.
- [6] J. Orozco, S. Campuzano, D. Kagan, M. Zhou, W. Gao, J. Wang, *Dynamic isolation and unloading of target proteins by aptamer-modified microtransporters*, *Analytical chemistry* 83(20) (2011) 7962-7969.
- [7] D. Kagan, S. Campuzano, S. Balasubramanian, F. Kuralay, G.-U. Flechsig, J. Wang, *Functionalized micromachines for selective and rapid isolation of nucleic acid targets from complex samples*, *Nano letters* 11(5) (2011) 2083-2087.
- [8] S. Balasubramanian, D. Kagan, C.M. Jack Hu, S. Campuzano, M.J. Lobo- Castañón, N. Lim, D.Y. Kang, M. Zimmerman, L. Zhang, J. Wang, *Micromachine- Enabled Capture and Isolation of Cancer Cells in Complex Media*, *Angewandte Chemie International Edition* 50(18) (2011) 4161-4164.
- [9] S. Campuzano, J. Orozco, D. Kagan, M. Guix, W. Gao, S. Sattayasamitsathit, J.C. Claussen, A. Merkoçi, J. Wang, *Bacterial isolation by lectin-modified microengines*, *Nano letters* 12(1) (2012) 396-401.
- [10] D. Vilela, J. Orozco, G. Cheng, S. Sattayasamitsathit, M. Galarnyk, C. Kan, J. Wang, A. Escarpa, *Multiplexed immunoassay based on micromotors and microscale tags*, *Lab on a Chip* 14(18) (2014) 3505-3509.
- [11] M. García, J. Orozco, M. Guix, W. Gao, S. Sattayasamitsathit, A. Escarpa, A. Merkoçi, J. Wang, *Micromotor-based lab-on-chip immunoassays*, *Nanoscale* 5(4) (2013) 1325-1331.
- [12] J. Wang, *Self-propelled affinity biosensors: Moving the receptor around the sample*, *Biosensors and Bioelectronics* 76 (2016) 234-242.
- [13] J. Orozco, B. Jurado-Sánchez, G. Wagner, W. Gao, R. Vazquez-Duhalt, S. Sattayasamitsathit, M. Galarnyk, A. Cortés, D. Saintillan, J. Wang, *Bubble-propelled micromotors for enhanced transport of passive tracers*, *Langmuir* 30(18) (2014) 5082-5087.
- [14] E. Morales- Narváez, M. Guix, M. Medina- Sánchez, C.C. Mayorga- Martinez, A. Merkoçi, *Micromotor enhanced microarray technology for protein detection*, *Small* 10(13) (2014) 2542-2548.
- [15] T. Ito, N. Aoki, S. Kaneko, K. Suzuki, *Highly sensitive and rapid sequential cortisol detection using twin sensor QCM*, *Analytical Methods* 6(18) (2014) 7469-7474.
- [16] P. Batista, A. Pereira, *Biomarkers in Neurodegenerative Diseases: Cortisol*, *J Mol Biomark Diagn* 7(277) (2016) 2.
- [17] B. Sun, Y. Gou, Y. Ma, X. Zheng, R. Bai, A.A.A. Abdelmoaty, F. Hu, *Investigate electrochemical immunosensor of cortisol based on gold nanoparticles/magnetic functionalized reduced graphene oxide*, *Biosensors and Bioelectronics* 88 (2017) 55-62.
- [18] R.M. Sapolsky, *Why stress is bad for your brain*, *Science* 273(5276) (1996) 749-751.

- [19] E.K. Do Yup Lee, M.H. Choi, Technical and clinical aspects of cortisol as a biochemical marker of chronic stress, *BMB reports* 48(4) (2015) 209.
- [20] N. Murase, S.-i. Taniguchi, E. Takano, Y. Kitayama, T. Takeuchi, A molecularly imprinted nanocavity-based fluorescence polarization assay platform for cortisol sensing, *Journal of Materials Chemistry B* 4(10) (2016) 1770-1777.
- [21] D. Appel, R.D. Schmid, C.-A. Dragan, M. Bureik, V.B. Urlacher, A fluorimetric assay for cortisol, *Analytical and bioanalytical chemistry* 383(2) (2005) 182-186.
- [22] T. Sink, R. Lochmann, K. Fecteau, Validation, use, and disadvantages of enzyme-linked immunosorbent assay kits for detection of cortisol in channel catfish, largemouth bass, red pacu, and golden shiners, *Fish Physiology and Biochemistry* 34(1) (2008) 95-101.
- [23] J. Fang, J.C. Zhou, E.H. Lan, B. Dunn, J.I. Zink, Bio-hybrid materials for immunoassay-based sensing of cortisol, *Journal of sol-gel science and technology* 50(2) (2009) 176-183.
- [24] S.K. Pasha, A. Kaushik, A. Vasudev, S.A. Snipes, S. Bhansali, Electrochemical immunosensing of saliva cortisol, *Journal of The Electrochemical Society* 161(2) (2014) B3077-B3082.
- [25] M. Moreno-Guzmán, M. Eguílaz, S. Campuzano, A. González-Cortés, P. Yáñez-Sedeño, J.M. Pingarrón, Disposable immunosensor for cortisol using functionalized magnetic particles, *Analyst* 135(8) (2010) 1926-1933.
- [26] A. Kaushik, A. Yndart, R.D. Jayant, V. Sagar, V. Atluri, S. Bhansali, M. Nair, Electrochemical sensing method for point-of-care cortisol detection in human immunodeficiency virus-infected patients, *International journal of nanomedicine* 10 (2015) 677.
- [27] A. Kaushik, A. Vasudev, S.K. Arya, S.K. Pasha, S. Bhansali, Recent advances in cortisol sensing technologies for point-of-care application, *Biosensors and Bioelectronics* 53 (2014) 499-512.
- [28] S. Campuzano, P. Yáñez-Sedeño, J.M. Pingarrón, Electrochemical bioaffinity sensors for salivary biomarkers detection, *TrAC Trends in Analytical Chemistry* 86 (2017) 14-24.




**Fig. 1.** Dynamic cortisol detection. (A) Schematic representation of dynamic cortisol-HRP detection using anti-cortisol-functionalized micromotors. (B) Micromotor-based cortisol-HRP naked-eye immunoassay detection system based on the use of TMB and  $H_2O_2$ . Absorbance spectra of samples prepared in the absence and presence of cortisol-HRP (light and dark blue lines, corresponding to transparent and blue color solutions, respectively).




**Fig. 2.** *Micromotor synthesis and characterization. (A) Micromotor synthesis and modification: template electrodeposition of PEDOT, Ni and Pt (a), release of the micromotors from the polycarbonate template (b), sputter of gold layer (c), and covalent antibody immobilization onto the Au-sputtered micromotors using a mixed MUA/MCH SAM and EDC/NHS chemistry (d). (B) SEM image of the PEDOT/Ni/Pt micromotor and corresponding EDX spectroscopy images illustrating the distribution of the Au, Ni and Pt layers. (C) Time-lapse images (8 s intervals) showing the propulsion and magnetic guidance of the antibody-modified PEDOT/Ni/Pt micromotors in cortisol spiked-buffer (images corresponding to video S1). Experimental conditions: PBS pH 7.4 solution containing 1%  $\text{H}_2\text{O}_2$ , 3% SDS and 1:100 of cortisol-HRP.*





**Fig. 3.** Influence of reaction time and reaction volume over the cortisol immunoassay detection. (A) Dependence of the interaction time between the anti-cortisol-micromotors (moving conditions) and anti-cortisol-microtubes (static conditions; in absence of inner Pt layer) with cortisol-HRP and control experiment performed with unmodified micromotors (no anti-cortisol immobilized). Experimental conditions: 50  $\mu$ L of 1:100 HRP-cortisol, 1%  $H_2O_2$ , 3% SDS, 500  $\mu$ L TMB solution, and  $1.1 \times 10^4$  micromotors. (B) Influence of the reaction volume (a-c: 25, 50 and 100  $\mu$ L, respectively) on the immunoreaction between HRP-cortisol and moving anti-cortisol-micromotors or static anti-cortisol-microtubes (red and purple lines, respectively). Experimental conditions: 2 min reaction, corresponding volume of 1:100 HRP-cortisol, 1%  $H_2O_2$ , 3% SDS, 500  $\mu$ L TMB solution, and  $1.1 \times 10^4$  micromotors.



**Fig. 4.** Comparison of direct competitive immunoassay for cortisol detection using static anti-cortisol-microtubes and moving anti-cortisol-micromotors. (A) UV-visible spectra corresponding to the oxidized TMB after performing the competitive micromotor-based immunoassay in (B) static and (C) moving conditions: in the absence and in the presence of  $0.1$  and  $1 \mu\text{g mL}^{-1}$  of free cortisol (a-c, respectively). Calibration plots for cortisol determination under static (D) and dynamic (E) conditions. Experimental conditions: 10 min reaction for static microtubes and 2 min for moving micromotors,  $50 \mu\text{L}$  of 1:100 HRP-cortisol, 1%  $\text{H}_2\text{O}_2$ , 3% SDS,  $500 \mu\text{L}$  TMB solution, and  $1.1 \times 10^4$  micromotors.